

PRODUCT INFORMATION

Product Name: CJC-1295 without DAC + IPA
 Batch / Lot No: PPA-RL-CP10-R00229
 Vial Size: 5 mg CJC-1295
 5 mg Ipamorelin
 Manufacturer: Peptaura-Retalux


TEST INFORMATION

Certificate No.: [PCB-CA1046]
 Sample Received: [07/10/2026]
 Analysis Date: [07/15/2026]
 Issue Date: [07/15/2026]
 Tested By: Pacific Coast Biologics Inc

TEST PARAMETER	METHOD	SPECIFICATION	RESULT	STATUS
Appearance	Visual	Lyophilized Powder	Lyophilized Powder	PASS
Endotoxin	LAL-Method	≤ 0.1 EU/μg	0.0444 EU/μg	PASS

CONCLUSION: Bacterial endotoxin testing was performed using the Limulus Amebocyte Lysate (LAL) method. The tested RUO-grade material demonstrated an endotoxin level of 0.0444 EU/μg, which meets the established acceptance criterion of ≤ 0.1 EU/μg. Therefore, the material **PASSED** the bacterial endotoxin specification.

Ramon Rochin
 Ramon Rochin
 Senior Chemist

Disclaimer: Results reported herein apply only to the sample(s) submitted for analysis. Testing is performed solely on the material received and does not constitute certification, approval, or endorsement of any product, batch, manufacturer, supplier, or intended use. No warranty, express or implied, is made regarding the safety, purity, efficacy, legality, merchantability, or fitness for a particular purpose of the tested material. The laboratory assumes no responsibility for the use, sale, distribution, labeling, marketing, or handling of any product associated with these results. Liability, if any, shall be limited to the fees paid for the testing services provided. This report may not be reproduced except in its entirety without prior written approval

***** ENDOSAFE Test Record *****
 PTS150 V11.1.2

DateTime: 15 Jul 2026 18:06:15
 Device: 21141224
 Analyst: RR
 Cartridge: Endotoxin-LAL
 Temperature: .. Start: 37.0C End: 37.0C
 Method: KX-122
 Cartridge Lot#: 2605541294
 Cartridge Cal Code: 119048543918
 Range: 1-0.01
 Range Time: Sec: 190-885
 Onset Time(s): 502 338 574 312
 Endotoxin Limit: 0.100 EU/ug
 Slope: -0.334 Intercept: +2.279
 Concentration: 1 ug/mL
 Sample Lot: 71526
 Sample Name: CP10
 Sample Rxn Time CV: 9.5% Pass
 Spike Value: 0.156 EU/mL
 Spike Rxn Time CV: 5.7% Pass
 Spike Recovery: 132% Pass
 Test Suitability: Pass
 Sample Endotoxin Limit: Pass
 Sample Value: 0.0444 EU/ug
 \$a6167a7fb865e41927c6d77734b21554aa8b5e38