

PRODUCT INFORMATION

Product Name: MOTS-C
Batch / Lot No: PPA-RL-MS40-R00227
Vial Size: 40 mg
Manufacturer: Peptaura-Retalux



TEST INFORMATION

Certificate No.: [PCB-CA1049]
Sample Received: [07/10/2026]
Analysis Date: [07/15/2026]
Issue Date: [07/15/2026]
Tested By: Pacific Coast Biologics Inc

TEST PARAMETER	METHOD	SPECIFICATION	RESULT	STATUS
Appearance	Visual	Lyophilized Powder	Lyophilized Powder	PASS
Endotoxin	LAL-Method	≤ 0.1 EU/μg	0.0243 EU/μg	PASS

CONCLUSION: Bacterial endotoxin testing was performed using the Limulus Amebocyte Lysate (LAL) method. The tested RUO-grade material demonstrated an endotoxin level of 0.0243 EU/μg, which did meet the established acceptance criterion of ≤ 0.1 EU/μg. Therefore, the material **PASSED** the bacterial endotoxin specification.

Ramon Rochin
Ramon Rochin
Senior Chemist

Disclaimer: Results reported herein apply only to the sample(s) submitted for analysis. Testing is performed solely on the material received and does not constitute certification, approval, or endorsement of any product, batch, manufacturer, supplier, or intended use. No warranty, express or implied, is made regarding the safety, purity, efficacy, legality, merchantability, or fitness for a particular purpose of the tested material. The laboratory assumes no responsibility for the use, sale, distribution, labeling, marketing, or handling of any product associated with these results. Liability, if any, shall be limited to the fees paid for the testing services provided. This report may not be reproduced except in its entirety without prior written approval

***** ENDOSAFE Test Record *****
PTS150 V11.1.2

DateTime: 15 Jul 2026 20:49:28
Device: 21141224
Analyst: RR
Cartridge: Endotoxin-LAL
Temperature: .. Start: 37.0C End: 37.0C
Method: KX-122
Cartridge Lot#: 2605541294
Cartridge Cal Code: 119048543918
Range: 1-0.01
Range Time: Sec: 190-885
Onset Time(s): 664 414 652 428
Endotoxin Limit: 0.100 EU/ug
Slope: -0.334 Intercept: +2.279
Concentration: 1 ug/mL
Sample Lot: 71526
Sample Name: MS40
Sample Rxn Time CV: 1.3% Pass
Spike Value: 0.068 EU/mL
Spike Rxn Time CV: 2.4% Pass
Spike Recovery: 58% Pass
Test Suitability: Pass
Sample Endotoxin Limit: Pass
Sample Value: 0.0243 EU/ug
\$5211fdc9c31c5c2294d4d50bc0905214ff5cfae5