

PRODUCT INFORMATION

Product Name: BPC-157
 Batch / Lot No: PPA-RL-BC10-R00224
 Vial Size: 10 mg
 Manufacturer: Peptaura-Retalux


TEST INFORMATION

Certificate No.: [PCB-CA1045]
 Sample Received: [07/10/2026]
 Analysis Date: [07/15/2026]
 Issue Date: [07/15/2026]
 Tested By: Pacific Coast Biologics Inc

TEST PARAMETER	METHOD	SPECIFICATION	RESULT	STATUS
Appearance	Visual	Lyophilized Powder	Lyophilized Powder	PASS
Endotoxin	LAL-Method	≤ 0.1 EU/μg	0.0216 EU/μg	PASS

CONCLUSION: Bacterial endotoxin testing was performed using the Limulus Amebocyte Lysate (LAL) method. The tested RUO-grade material demonstrated an endotoxin level of 0.0216 EU/μg, which meets the established acceptance criterion of ≤ 0.1 EU/μg. Therefore, the material **PASSED** the bacterial endotoxin specification.

Ramon Rochin
 Ramon Rochin
 Senior Chemist

Disclaimer: Results reported herein apply only to the sample(s) submitted for analysis. Testing is performed solely on the material received and does not constitute certification, approval, or endorsement of any product, batch, manufacturer, supplier, or intended use. No warranty, express or implied, is made regarding the safety, purity, efficacy, legality, merchantability, or fitness for a particular purpose of the tested material. The laboratory assumes no responsibility for the use, sale, distribution, labeling, marketing, or handling of any product associated with these results. Liability, if any, shall be limited to the fees paid for the testing services provided. This report may not be reproduced except in its entirety without prior written approval

***** ENDOSAFE Test Record *****

PTS150 V11.1.2

DateTime: 15 Jul 2026 17:28:55

Device: 21141224

Analyst: RR

Cartridge: Endotoxin-LAL

Temperature: .. Start: 37.0C End: 37.0C

Method: KX-122

Cartridge Lot#: 2605541294

Cartridge Cal Code: 119048543918

Range: 1-0.01

Range Time: Sec: 190-885

Onset Time(s): 694 390 674 342

Endotoxin Limit: 0.100 EU/ug

Slope: -0.334 Intercept: +2.279

Concentration: 1 ug/mL

Sample Lot: 71526

Sample Name: BC10

Sample Rxn Time CV: 2.1% Pass

Spike Value: 0.119 EU/mL

Spike Rxn Time CV: 9.3% Pass

Spike Recovery: 101% Pass

Test Suitability: Pass

Sample Endotoxin Limit: Pass

Sample Value: 0.0216 EU/ug

\$f63da2066564fe99a865323733949668c030e43f